

14 Amp Laser Diode Driver, 24 Volt Compliance Range



14 Amp, 24 Volt Laser Diode Driver High Compliance Voltage Range

- o Current up to 14 A, Voltage up to 24 V
- o Optimized for High Compliance Voltage Pump Laser Diodes from nLight, Lumics, II-VI, Lumentum, Coherent/Dilas
- o CW Mode and Integrated Quasi-CW Pulse Generator; Pulse Widths from 30 μ s to CW
- o User-Programmable Soft-Start Current Ramp to Laser Diode Current Setpoint
- o Open Circuit Detection and Fast Shut-Down with Analog Control Loop





**LASER
DIODE
DRIVERS**

The LDI-127 high-compliance laser diode driver is designed to precisely and safely bias high-power a wide range of multi-single emitter lasers, such as nLight, Lumentum and II-IV. The LDI-127 also drives multi-emitter devices wired for series operation.

The LDI-127 operates in CW (continuous wave) mode, and also provides flexible modulation capabilities and a QCW mode. On the backpanel is the BNC input for an analog or TTL digital modulation, DC up to 50 kHz (10k Ω input impedance).

The controller has an internal function generator which can be used to drive quasi-CW pulses in continuous, single, and burst-mode. In QCW mode, the user can also set 25 μ s to CW pulses to trigger from a remote TTL signal source.





LDI-127 High Power Laser Diode Driver Specifications

LASER DIODE CURRENT AND VOLTAGE OUTPUT

- Compliance Voltage Range: 0.00 - 24.00 Volts
- Output Current Range: 0.00 - 14.00 Amps
- Current Noise & Ripple: 1% to 0.01% of I_{max} (rms)
- Current Setpoint Resolution: 3.5 mA
- Current Setpoint Accuracy: $\pm 0.5\%$
- Current Stability (4 hours): ≤ 100 ppm
- Current Limit Setpoint Accuracy: $\pm 2\%$

INTEGRATED LASER DIODE PROTECTION FEATURES

- Programmable Soft-Start Current Ramp to Set Point (300ms Default)
- Independent Pilot Laser Output (5V, 150mA max)
- User-Programmable Current Limit
- Open Circuit Detection
- ESD and Power Surge Clamp, AC Line Filter
- Reverse Voltage Transient Clamp
- Rear Panel Keylock Switch and Safety Interlock
- Short Circuit when Laser Diode Current Turned OFF

INTERNAL PULSE MODE

- Pulse Width Range: (model specific) 10 μ s - 71 minutes
- Pulse Period Range: Pulse Width + 100 μ s
- Time Base Accuracy: $\pm 1\%$
- Pulse-to-Pulse Accuracy: 30 μ s
- Pulse Rise Time: ~ 18 μ s
- Pulse Fall Time: 70% - 100% of Rise Time



LDI-127 High Power Laser Diode Driver Specifications

MODULATION AND QCW MODE

- QCW Mode Rise / Fall Time: 20 μ s to CW, 10%-90% (< 5 μ s on request)
- QCW Trigger: Internal Pulse Generator or External
- Modulation Input (BNC): Digital (TTL) or Analog
- Modulation BNC Input Impedance: 10K
- -----
- Modulation Input Voltage Range: 0 ~ 4 Volts (4V = Max Current)
- Analog Modulation Bandwidth: 1 Hz – 20 kHz

AUXILIARY FUNCTIONS

- Temperature Sensor Input: 10k Ω NTC Thermistor
- External Fan Control Circuit, 1 - 24V, 300mA (max)
- Laser-On External LED Indicator: 5mA Output
- Pilot Laser Anode, vs. Ground
- Photodiode Current Measurement Accuracy: \pm 0.5%
- Photodiode Current Measurement Range: 0.00 - 700 μ A

USER INTERFACE

- Front Panel: Alphanumeric LCD
- RS232 Standard
- USB Optional: \$95.00 (Option SVC-USB)
- LabView Drivers Included

DIMENSIONS AND POWER INPUT

- Universal 110V ~ 240 VAC Input
- 95mm (H) x 119mm (W) x 280mm (L)

RECOMMENDED ACCESSORIES

- kab-39 Unterminated Connecting Cable -or- kab-231 Terminated Connecting Cable
- kab-286 Unterminated Power Cable -or- kab-297 Terminated Power Cable
- acc-417 USB-RS232 Converter

Product Sales and Service

Orders for this product are fulfilled by LaserDiodeControl.com, part of the Laser Lab Source group. It is manufactured for Laser Lab Source by OsTech, GmbH.

Product Warranty

This product is sold with a full one-year warranty. It is warrantied to be free from defects in material and/or workmanship for a period of one year from the date of shipment.



Laser Lab Source
670 S. Ferguson St., Suite 3
Bozeman, MT 59718 USA
800-887-5065
LaserLabSource.com

Ostech, GmbH
Plauener Str. 163-165 • Haus i • 13053
Berlin